| Model Number 137B27 | ICP [®] PRESSURE SENSOR | | | SENSOR | Revision: A ECN #: 51260 |
|---|---|---------------------------|------------|---|--------------------------------------|
| erformance | ENGLISH | SI | | OPTIONAL VERSIONS | e |
| leasurement Range(for ±5V output) | 500 psi | 3,447 kPa | | Optional versions have identical specifications and accessories a where noted below. More than one option | s listed for the standard model exce |
| eful Overrange(for ± 10V output) | 1,000 psi | 6,895 kPa | [1] | where noted below. More than one option | n may be used. |
| nsitivity($\pm 15\%$) | 10 mV/psi | 1.45 mV/kPa | [2] | | |
| ximum Pressure | | 34,474 kPa | [2] | | |
| olution | 5 kpsi | | [2] | | |
| | 1 mpsi | 0.007 kPa | [3] | | |
| onant Frequency | ≥ 400 kHz | ≥ 400 kHz | | | |
| e Time(Incident) | ≤ 6.5 µ sec | ≤ 6.5 µ sec | | | |
| n-Linearity | ≤ 1.0 % FS | ≤ 1.0 % FS | [4] | | |
| vironmental | | | | | |
| nperature Range(Operating) | -100 to +275 °F | -73 to +135 ℃ | | | |
| nperature Coefficient of Sensitivity | ≤ 0.05 %/°F | ≤ 0.090 %/°C | | | |
| ctrical | | | | | |
| charge Time Constant(at room temp) | ≥ 0.2 sec | ≥ 0.2 sec | | | |
| itation Voltage | 20 to 30 VDC | 20 to 30 VDC | | | |
| istant Current Excitation | 2 to 20 mA | 2 to 20 mA | | | |
| | | | | | |
| put Impedance | ≤ 100 Ohm | ≤ 100 Ohm | | | |
| put Bias Voltage | 8 to 14 VDC | 8 to 14 VDC | | | |
| /sical | | _ · | | | |
| sing Geometry | Compression | Compression | | | |
| sing Element | Quartz | Quartz | | | |
| using Material | Aluminum | Aluminum | | | |
| phragm | Invar | Invar | | | |
| ling | Ероху | Epoxy | | | |
| trical Connector | 4-Pin | 4-Pin | [2] | | |
| ight | 16.1 oz | 456 gm | [3] | | |
| | | | | NOTES: [1]For +10 volt output, minimum 26 VDC supply voltage required limited by output bias. [2]Two identically spec'd elements spaced 10cm apart - see dra calibration certs supplied for each channel. [3]Typical. [4]Zero-based, least-squares, straight line method. [5]See PCB Declaration of Conformance PS023 for details. SUPPLIED ACCESSORIES: Model PCS-6 Calibration of Series 134, 137 (2-sensor probes onl | wing #65310 for details. Individua |
| . C ^[5] | | | | Entered: LK Engineer: RPF Sales: RWM Ap | proved: RPF Spec Number: |
| | | | | Date: 11/13/2020 Date: 11/13/2020 Date: 11/13/2020 Da | |
| pecifications are at room temperature unl | less otherwise specified. | | | WPCB PIEZOTRONICS 3425 Walden Avenue, Depew, NY 14043 | 0987 |
| e interest of constant product improveme | nt, we reserve the right to ch onics, Inc. | ange specifications witho | ut notice. | | |