
Model Number						
356 \ \ 71						

TRIAXIAL CHARGE OUTPUT ACCELEROMETER

Revision: D ECN #: 40410

Performance	ENGLISH	<u>SI</u>	
Sensitivity(± 30 %)	10 pC/g	1.02 pC/(m/s ²)	
Measurement Range	± 500 g pk	± 4905 m/s² pk	
Frequency Range(+5 %)	5 kHz	5 kHz	[2]
Frequency Range(+10 %)	7 kHz	7 kHz	[2]
Resonant Frequency	≥ 25 kHz	≥ 25 kHz	
Non-Linearity	≤ 1 %		[3]
Transverse Sensitivity	≤ 5 %	≤ 5 %	
Environmental			
Overload Limit(Shock)	± 5000 g pk	± 49,000 m/s ² pk	
Temperature Range	-95 to +490 °F	-70 to +254 °C	
Temperature Response	See Graph	See Graph	[1]
Electrical			
Capacitance	690 pF	690 pF	[1]
Insulation Resistance(at 70 °F [21 °C])	>10 ¹² Ohm	>10 ¹² Ohm	
Insulation Resistance(at 490 °F [254 °C])	>10 ⁸ Ohm	>10 ⁸ Ohm	[1]
Output Polarity	Negative	Negative	
Physical			
Sensing Element	Ceramic	Ceramic	
Sensing Geometry	Shear	Shear	
Housing Material	Titanium	Titanium	
Sealing	Hermetic	Hermetic	
Size (Length x Width x Height)	0.96 in x 1.00 in x 0.50 in	24.4 mm x 25.4 mm x 12.7 mm	
Weight	0.8 oz	22.7 gm	[1]
Electrical Connector	10-32 Coaxial Jack	10-32 Coaxial Jack	
Mounting	Through Hole	Through Hole	
Mounting Torque	8 to 10 in-lb	90 to 113 N-cm	

Typical Sensitivity Deviation vs Temperature



All specifications are at room temperature unless otherwise specified.

In the interest of constant product improvement, we reserve the right to change specifications without notice.

ICP® is a registered trademark of PCB Group, Inc.

OPTIONAL VERSIONS

Optional versions have identical specifications and accessories as listed for the standard model except where noted below. More than one option may be used.

M - Metric Mount

Supplied Accessory : Model M039A22 Allen wrench for use with M3 thread (1) replaces Model 039A22

Supplied Accessory : Model M080A170 Adhesive mounting base (1) replaces Model 080A170 Supplied Accessory : Model M081A94 Mounting Stud, M3 X 0.50 X 20mm (2) replaces Model 081A94

P - Positive Output Polarity

NOTES:

- [1] Typical.
- [2] Low frequency response is determined by external signal conditioning electronics.
- [3] Zero-based, least-squares, straight line method.

SUPPLIED ACCESSORIES:

Model 039A22 Allen wrench, 7/64" hex (1)

Model 080A170 Adhesive mounting base (for Model 356A71) (1)

Model 080A90 Quick Bonding Gel (1)

Model 081A94 Mounting stud, 6-32 UNC, 0.75" long (2)

Model ACS-1T NIST traceable triaxial amplitude response, 10 Hz to upper 5% frequency. (1)

Entered:	Engineer: MAM	Sales: WDC	Approved: BAM	Spec Number:
Date:	Date: 1/29/2013	Date: 1/29/2013	Date: 1/29/2013	15360



Phone: 716-684-0001 Fax: 716-684-0987 E-Mail: info@pcb.com