

Model 482C05

Four-channel, ICP® Sensor Signal Conditioner

Installation and Operating Manual

For assistance with the operation of this product, contact PCB Piezotronics, Inc.

Toll-free: 800-828-8840 24-hour SensorLine: 716-684-0001 Fax: 716-684-0987 E-mail: info@pcb.com Web: www.pcb.com







Repair and Maintenance

PCB guarantees Total Customer Satisfaction through its "Lifetime Warranty Plus" on all Platinum Stock Products sold by PCB and through its limited warranties on all other PCB Stock, Standard and Special products. Due to the sophisticated nature of our sensors and associated instrumentation, field servicing and repair is not recommended and, if attempted, will void the factory warranty.

Beyond routine calibration and battery replacements where applicable, our products require no user maintenance. Clean electrical connectors, housings, and mounting surfaces with solutions and techniques that will not harm the material of construction. Observe caution when using liquids near devices that are not hermetically sealed. Such devices should only be wiped with a dampened cloth—never saturated or submerged.

In the event that equipment becomes damaged or ceases to operate, our Application Engineers are here to support your troubleshooting efforts 24 hours a day, 7 days a week. Call or email with model and serial number as well as a brief description of the problem.

Calibration

Routine calibration of sensors and associated instrumentation is necessary to maintain measurement accuracy. We recommend calibrating on an annual basis, after exposure to any extreme environmental influence, or prior to any critical test.

PCB Piezotronics is an ISO-9001 certified company whose calibration services are accredited by A2LA to ISO/IEC 17025, with full traceability to SI through N.I.S.T. In addition to our standard calibration services, we also offer specialized tests, including: sensitivity at elevated or cryogenic temperatures, phase response, extended high or low frequency response, extended range, leak testing, hydrostatic pressure testing, and others. For more information, contact your local PCB Piezotronics distributor, sales representative, or factory customer service representative.

Returning Equipment

If factory repair is required, our representatives will provide you with a Return Material Authorization (RMA) number, which we use to reference any information you have already provided and expedite the repair process. This number should be clearly marked on the outside of all returned package(s) and on any packing list(s) accompanying the shipment.

Contact Information

PCB Piezotronics, Inc. 3425 Walden Ave. Depew, NY14043 USA Toll-free: (800) 828-8840 24-hour SensorLine: (716) 684-0001 General inquiries: <u>info@pcb.com</u> Repair inquiries: <u>rma@pcb.com</u>

For a complete list of distributors, global offices and sales representatives, visit our website, <u>www.pcb.com</u>.

Safety Considerations

This product is intended for use by qualified personnel who recognize shock hazards and are familiar with the precautions required to avoid injury. While our equipment is designed with user safety in mind, the protection provided by the equipment may be impaired if equipment is used in a manner not specified by this manual.

Discontinue use and contact our 24-Hour Sensorline if:

- Assistance is needed to safely operate equipment
- Damage is visible or suspected
- Equipment fails or malfunctions

For complete equipment ratings, refer to the enclosed specification sheet for your product.

Definition of Terms and Symbols

The following symbols may be used in this manual:



DANGER

Indicates an immediate hazardous situation, which, if not avoided, may result in death or serious injury.



CAUTION

Refers to hazards that could damage the instrument.



NOTE

Indicates tips, recommendations and important information. The notes simplify processes and contain additional information on particular operating steps.

The following symbols may be found on the equipment described in this manual:



This symbol on the unit indicates that high voltage may be present. Use standard safety precautions to avoid personal contact with this voltage.



This symbol on the unit indicates that the user should refer to the operating instructions located in the manual.



This symbol indicates safety, earth ground.



PCB工业监视和测量设备 - 中国RoHS2公布表 PCB Industrial Monitoring and Measuring Equipment - China RoHS 2 Disclosure Table

	有害物 质								
部件名称	铅 (Pb)	汞 (Hg)	镉 (Cd)	六价铬 (Cr(VI))	多溴 联苯 (PBB)	多溴二苯 醚 (PBDE)			
住房	0	0	0	0	0	0			
PCB板	Х	0	0	0	0	0			
电气连接 器	0	0	0	0	0	0			
压电晶 体	х	0	0	0	0	0			
环氧	0	0	0	0	0	0			
铁氟龙	0	0	0	0	0	0			
电子	0	0	0	0	0	0			
厚膜基板	0	0	Х	0	0	0			
电线	0	0	0	0	0	0			
电缆	Х	0	0	0	0	0			
塑料	0	0	0	0	0	0			
焊接	Х	0	0	0	0	0			
铜合金 /黄 铜	Х	0	0	0	0	0			
本表格依据 SJ/T 11364 的规定编制。									
O:表示该有害物质在该部件所有均质材料中的含量均在 GB/T 26572 规定的限量要求以下。									
X:表示该有害物质至少在该部件的某一均质材料中的含量超出 GB/T 26572 规定的限量要求。 铅是欧洲RoHS指令2011/65/ EU附件三和附件四目前由于允许的豁免。									

CHINA ROHS COMPLIANCE

Component Name	Hazardous Substances								
	Lead (Pb)	Mercury (Hg)	Cadmium (Cd)	Chromium VI Compounds (Cr(VI))	Polybrominated Biphenyls (PBB)	Polybrominated Diphenyl Ethers (PBDE)			
Housing	0	0	0	0	0	0			
PCB Board	Х	0	0	0	0	0			
Electrical Connectors	0	0	0	0	0	0			
Piezoelectric Crystals	Х	0	0	0	0	0			
Ероху	0	0	0	0	0	0			
Teflon	0	0	0	0	0	0			
Electronics	0	0	0	0	0	0			
Thick Film Substrate	0	0	Х	0	0	0			
Wires	0	0	0	0	0	0			
Cables	Х	0	0	0	0	0			
Plastic	0	0	0	0	0	0			
Solder	Х	0	0	0	0	0			
Copper Alloy/Brass	Х	0	0	0	0	0			

This table is prepared in accordance with the provisions of SJ/T 11364.

O: Indicates that said hazardous substance contained in all of the homogeneous materials for this part is below the limit requirement of GB/T 26572.

X: Indicates that said hazardous substance contained in at least one of the homogeneous materials for this part is above the limit requirement of GB/T 26572.

Lead is present due to allowed exemption in Annex III or Annex IV of the European RoHS Directive 2011/65/EU.



SIGNAL CONDITIONER

Models 482C05/482C15

GENERAL OPERATION MANUAL

For powering ICP[®] sensors and in-line charge converters, this signal conditioner provides a cost effective method for powering small numbers of sensor channels.

The 482C Series is available with many different options from PCB. This manual describes the functions of the most economical signal conditioner with unity gain (482C05) and x1, x10, x100 jumper selectable gain (482C15), of that series.



MANUAL #: 38410 MANUAL REV: B ECO: 50033 10/31/2019

PCB PIEZOTRONICS, INC. 3425 WALDEN AVENUE DEPEW, NY 14043-2495 PHONE 716-684-0001 FAX 716-684-0987

[®]ICP is a registered trademark of PCB Group, Inc.

KS398

Table of Contents

Table of	of Contents	i
1-1.	Introduction	1
1-2.	Safety Considerations	1
1-3.	Equipment Rating	2
1-4.	Model 482C15 System Description	2
1-5.	Model 482C05 System Description	4
1-6.	Installation	4
1-7.	Grounding Techniques	4
1-8.	Unit Power	5
1-9.	Maintenance and Repair	5

Appendix Specification Sheet Outline Drawing

1-1. Introduction

The Model 482C05 is a basic ICP signal conditioner that provides an adjustable current source to drive ICP[®] sensors. Additionally it has Open, Short and Overload Input fault condition indicators.

The Model 482C15 is an ICP[®] signal conditioner that has all the features of the 482C05 plus jumper selectable (x1, x10, and x100) Gain options and an Overload range selection. An easy to open cover, allows access to the jumper selections.

The following sections describe these devices in more detail.

1-2. Safety Considerations

WARNING SYMBOLS AND TERMS

The following symbols and terms may be found on the equipment described in this manual.

This symbol on the unit indicates that the user should refer to the operating instructions located in the manual.

This symbol on the unit indicates that high voltage may be present. Use standard safety precautions to avoid personal contact with this voltage.

This symbol indicates that the test fixture, Model 482C15, must be connected to earth ground via the power cord.

The **WARNING** heading used in this manual explains dangers that might result in personal injury or death. Always read the associated information very carefully before performing the indicated procedure.

The CAUTION heading used in this manual explains hazards that could damage the instrument.

WARNING 1: The power supply/signal conditioner should not be opened by anyone other than qualified service personnel. This product is intended for service by qualified personnel who recognize shock hazards and are familiar with the safety precautions required to avoid injury.

WARNING 2: This equipment is designed with user safety in mind; however, the protection provided by the equipment may be impaired if the equipment is used in a manner not specified by PCB Piezotronics, Inc.

CAUTION 1: *Cables can kill your equipment.* High voltage <u>ElectroStatic Discharge (ESD)</u> can damage electrical devices. Similar to a capacitor, a cable can hold a charge caused by triboelectric transfer, such as that which occurs in the following:

- Laying on and moving across a rug.
- Any movement through air.
- The action of rolling out a cable.
- Contact to a non-grounded person

The solution for product safety: 1) Connect the cables only with the AC power off. 2) Temporarily "short" the end of the cable before attaching it to any signal input or output.

1

CAUTION 2: *ESD considerations should be made prior to performing any internal adjustments on the equipment.* Any piece of electronic equipment is vulnerable to ESD when opened for adjustments. Internal adjustments should therefore be done ONLY at an ESD-safe work area. Many products have ESD protection, but the level of protection may be exceeded by extremely high voltage that is typically present in normal situations.

1-3. Equipment Rating

For complete specifications, please refer to the enclosed Specification Sheet. This equipment operates optimally at +32 to $+120^{\circ}$ F (0 to $+50^{\circ}$ C), in an environment having <85% relative humidity. Its line power frequency range is 47-63 Hz.

The 482C15/482C05 devices require 34-38 VDC with 200 mA to operate. In turn, they supply 26 VDC @ 2 to 20 mA excitation voltage to connected ICP®sensors or in-line charge converters. The units get power via AC power adaptor PCB Model 488A04, 36VDC battery pack, or a car adapter which all plug into a receptacle on the rear panel (PCB Model 488A11).

1-4. Model 482C15 System Description

The Model 482C15 is a four channel signal conditioning unit designed to interface measurement signals to readout or recording devices.

The following are features that come standard on 482C15 units:

- Input fault and overload LED's with jumper selectable overload level of +/-5V or +/-10V
- Variable ICP[®] source current, 0 to 20 mA
- Jumper selectable Voltage input mode
- Jumper selectable gain of x1, x10, x100 (482C15 Only)
- Lighted power indicator on the front panel
- Rear panel BNC input/output connectors
- Optional L.P. plug in filters (477A series type) various 2-pole frequencies avail. Consult factory

1-4.1. Model 482C15 ICP[®] Input/Output Mode

The Model 482C15 contains a regulated 26 VDC power supply that provides constant current for up to 4 individual channels. Both the output and input connections utilize BNC connectors and are brought out through the rear panel. The unit is AC coupled only. The 482C series offers optional DC Coupling. Consult factory

1-4.2. Adjusting the Current Source

The Current adjustment is located on the rear of the unit. Use a standard small flat screwdriver to adjust the "I ADJ" potentiometer. Turning the adjustment slot clockwise will increase the current, while counterclockwise will decrease the current. Do this adjustment only while a DC mA meter or DMM set to mA is connected to any "ICP" BNC. One adjustment controls all channels. The unit is factory set to 4mA and can be adjusted between 0mA-20mA.

1-4.3. Fault Indications

The three red LED's on the front panel are all indicators of some type of problem. The normal good operation state is when all fault lights are off. The "OPEN" light indicates that the Bias voltage is over that of most sensors and is close to an open circuit voltage of 26VDC. The photo on the cover is shown with the unit on, and no sensors attached, so the "OPEN" LED is on. The "SHORT" light indicates a very low Bias voltage close to 0VDC. The "OVERLOAD" light


2

indicates a signal level of +/-10V Peak has been detected and the signal may be in danger of clipping. These fault lights are active for all channels, at the same time.

1-4.4. Gain selection / Jumper definition

The 482C15 has internal jumpers that are accessed by opening the top cover. The top cover is secured by four (4) Philips head screws on the bottom of the unit. Place the unit on a clean, flat surface (Electrical Static Discharge protection recommended) and remove the four screws. Carefully rotate the unit and place it on its feet, then gently remove the cover.

The various jumpers are all labeled on the PC board. Each channel can be set to your selected gain and other options. The selection of all these features with jumpers allows the 482C15 to have the functionality of many other higher priced units at an economical cost. See the table that explains the function and number of each jumper. "X" represents channel number. Once jumpers have been placed as needed, replace cover – taking special care to align the front and back panels with the slots in the top cover. Insert the four (4) screws and tighten to 5 in-lbs.



Gain 1 10 100 Overload 5V/10V	JX02 J1	1 jumper per channel Controls all channels
Voltage Mode	JX03	2 jumpers per channel
Bypass Filter	JX00	Remove to install filter module
477A Filter	FILX00	Install filter here per channel

Model 482C05 System Description

The Model 482C05 is a basic four channel signal conditioning unit designed to interface measurement signals to readout or recording devices.

The following are features that come standard on every 482C05:

- Unity gain
- Input fault and overload LED's
- Variable ICP® source current, 0 to 20 mA
- Lighted power indicator on the front panel
- Rear panel BNC input/output connectors

1-4.5. Model 482C05 ICP[®] Input/Output Mode

The Model 482C05 contains a regulated 26 VDC power supply that provides constant current for up to 4 individual channels. Both the output and input connections utilize BNC connectors and are brought out through the rear panel. The unit is AC coupled only. The 482C series offers optional DC Coupling. Consult factory

1-4.6. Adjusting the Current Source

The Current adjustment is located on the rear of the unit. Use a standard small flat screwdriver to adjust the "I ADJ" potentiometer. Turning the adjustment slot clockwise will increase the current, while counterclockwise will decrease the current. Do this adjustment only while a DC mA meter or DMM set to mA is connected to any "ICP" BNC. One adjustment controls all channels. The unit is factory set to 4mA and can be adjusted between 0mA-20mA.

1-4.7. Fault Indications

The three red LED's on the front panel are all indicators of some type of problem. The normal good operation state is when all fault lights are off. The "OPEN" light indicates that the Bias voltage is over that of most sensors and is close to an open circuit voltage of 26VDC. The photo on the cover is shown with the unit on, and no sensors attached, so the "OPEN" LED is on. The "SHORT" light indicates a very low Bias voltage close to 0VDC. The "OVERLOAD" light indicates a signal level of +/-10V Peak has been detected and the signal may be in danger of clipping. These fault lights are active for all channels, at the same time.

1-5. Installation

Both the Model 482C15 and the Model 482C05 come packaged in a standard box and require very little in term of installation. The box should be located in such a way as to allow convenient access to the power outlet for disconnect purposes. Since these units have low power consumption, they can be located in confined environments.

NOTE: Specific installation is available; contact PCB

1-6. Grounding Techniques

Integrating either of these Models into an application that links the outputs to other test equipment powered by line voltages may lead to errors or loss of signal-to-noise ratio due to ground loops. The evidence of ground loops is easily seen whenever the fundamental frequency (50 or 60 Hz) or a multiple of the fundamental frequency is present in the system when the sensors are "at rest." In order to maintain the operating specification of noise and reduce the effects of line interference, proper grounding techniques should be used. The following procedure may be helpful:

1. Make sure the signal ground lines of all equipment are tied together. The signal grounds of the channels are typically tied together via the case of the input and output BNC connectors. The individual channels of the Model 482C15 and 482C05 have their signal ground line tied together internally at the power supply.

4

2. Insure that the sensor does not pick up line noise from the body under test. The case of the sensor should be isolated from the body under (ground) using an isolation pad. The isolation pad breaks the loop formed by the signal path of the sensor to the unit and the return (ground) from the unit back to the sensor.

3. Make sure that all equipment signal grounds of the test system are tied to the Earth ground at a single point and the connections linking the equipment ground are made using a wire that can provide a very low impedance connection.

1-7. Unit Power

1-7.1. Power: Standard AC Line adapter Model 488A04 (supplied with unit)

Plug the adapter into any 100-240VAC 47-63Hz power source, attach the supplied ferrite clamp on the AC cord to ensure CE conformance, and press the "ON/OFF" button.

1-7.2. Power: Optional Auto adapter Model 488A11

Plug the adapter into a 12VDC power source, or Automobile cigar lighter socket and press the "ON/OFF" button.

1-7.3. Power: Using other 36VDC Power Sources, or batteries. (Not supplied with unit)


When using any 36VDC power source to power the unit, make sure to observe the correct pin out, as indicated on the rear panel of the unit. Standard 5-pin din connector.

CAUTION 3: Refer to the rear panel for proper input voltage. Only PCB supplied power sources will ensure noise specifications and CE ratings.

1-8. Maintenance and Repair

It is not recommended that the customer attempt repairs of the Model 482C15 or 482C05. Should trouble occur, contact the factory for assistance.

Model Number 482C05	FO	UR-CHANNEI	_, ICP® SEN	SOR	SIGNAL	CONDITI	ONER		ision: J I #: 43617	
Performance		ENGLISH	<u>SI</u> 4			ОР [.]	TIONAL VERS			
Channels		4	-		Optional versions have identical specifications and accessories as listed for the standard					
Sensor Input Type(s)		ICP®	ICP®		model except where noted below. More than one option may be used.					
Output Range(Maximum	/	± 10 V	± 10 V							
Low Frequency Respons		<0.1 Hz	<0.1 Hz	[2][3]						
High Frequency Respons		>1000 kHz ± 1 °	>1000 kHz ± 1 °							
Phase Response(at 1 kH	12)	= :	= -							
Cross Talk(maximum)		-72 dB	-72 dB							
Fault/Bias Monitor/Meter	(LED)	Open/Short/Overload	Open/Short/Overload							
Environmental		100 to 1400 %F	0.4- 1.50.80							
Temperature Range(Ope	erating)	+32 to +120 °F	0 to +50 °C							
Electrical										
Power Required(for supp	• • • •	AC Power	AC Power		NOTES:					
Power Required(direct in	iput to unit)	DC power	DC power		[1]User adjustabl	e, factory set at 4 m	A (± 0.5 mA). One c	control adjusts all cha	nnels.	
AC Power(47 to 63 Hz)		100 to 240 VAC	100 to 240 VAC			n input impedance o				
AC Power		≤ 0.7 Amps	≤ 0.7 Amps				e input impedance a	affects discharge time	e constant and low	
Excitation Voltage(± 1 VI	DC)(To Sensor)	+26 VDC	+26 VDC		frequency response of unit.					
DC Offset		<20 mV	<20 mV		[4]Typical.					
DC Power		+32 to 38 VDC	+32 to 38 VDC		[5]See PCB Declaration of Conformance PS024 for details.					
DC Power		<0.25 Amps	<0.25 Amps							
Constant Current Excitat Overload Threshold(± 1.0		2 to 20 mA ± 10 Vpk	2 to 20 mA ± 10 Vpk	[1]						
Discharge Time Constan	1 /	± 10 vpk >5 sec	± 10 νρκ >5 sec	101[01						
Broadband Electrical Noi		3.5 µV rms	3.5 µV rms	[2][3] [4]						
Spectral Noise(1 Hz)	130(11010,000112)	1.30 µV/√Hz	1.30 µV/√Hz	[4]						
Spectral Noise(10 Hz)		0.10 µV/√Hz	0.10 µV/√Hz	[4]						
Spectral Noise(100 Hz)		0.08 µV/√Hz	0.08 µV/√Hz	[4]						
Spectral Noise(1 kHz)		0.07 µV/√Hz	0.07 µV/√Hz	[4] [4]						
Spectral Noise(10 kHz)		0.07 µV/√Hz	0.07 µV/√Hz	[4]						
Physical Electrical Connector(ICP		BNC Jack	BNC Jack		SUPPLIED ACCESSORIES:					
(1 /				Model 017AXX Power Cord					
Electrical Connector(Out	. ,	BNC Jack	BNC Jack		Model 488B04/N	C Power Convertor				
Electrical Connector(DC Size (Height x Width x D		5-socket DIN (female) 3.2 in x 8.0 in x 5.9 in	5-socket DIN (female) 8.1 cm x 20 cm x 15 cm							
Size (Height x Width x D	epui)	3.2 in x 8.0 in x 5.9 in 1.25 lb								
weight		1.25 10	567 gm		Entered: AP	Engineer: CPH	Sales: ML	Approved: JWH	Spec Number:	
"					Date: 1/28/2015	Date: 1/28/2015	Date: 1/28/2015	Date: 1/28/2015	35061	
In the interest of constan	oom temperature unless othe t product improvement, we re emark of PCB Group, Inc.	erwise specified. eserve the right to change sp	ecifications without notice.			B PIEZOTH enue, Depew, NY 14		Fax: 716-6	6-684-0001 84-0987 o@pcb.com	



3

	1	
	REVISIONS	
REV	DESCRIPTION	DIN
Е	UPDATED ARTWORK	37737

В